

BC-6D

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT

 MB0326A

2026-05-10

Instrument	Parameter		Low		Normal		High	
			<div>LOT</div>	MB0326AL	<div>LOT</div>	MB0326AN	<div>LOT</div>	MB0326AH
BC-7500[B] CRP	WBC	$\times 10^9/L$	3.71	± 0.50	7.90	± 1.00	21.17	± 2.50
BC-7500[N] CRP	Neu#	$\times 10^9/L$	2.08	± 0.37	4.00	± 0.80	11.44	± 2.12
BC-7500[R] CRP	Lym#	$\times 10^9/L$	0.92	± 0.30	2.71	± 0.64	7.13	± 1.70
BC-7500[NR] CRP	Mon#	$\times 10^9/L$	0.13	± 0.13	0.38	± 0.38	0.99	± 0.99
BC-7300[B] CRP	Eos#	$\times 10^9/L$	0.52	± 0.34	0.70	± 0.70	1.31	± 1.31
BC-7300[R] CRP	Bas#	$\times 10^9/L$	0.06	± 0.20	0.11	± 0.40	0.30	± 1.08
BC-7300[NR] CRP	Neu%	%	56.4	± 10.0	50.6	± 10.0	54.0	± 10.0
BC-7500[B] CS	Lym%	%	24.7	± 8.0	34.3	± 8.0	33.7	± 8.0
BC-7500[N] CS	Mon%	%	3.5	± 3.5	4.8	± 4.8	4.7	± 4.7
BC-7500[R] CS	Eos%	%	13.9	± 9.0	8.9	± 8.9	6.2	± 6.2
BC-7500[NR] CS	Bas%	%	1.5	± 5.0	1.4	± 5.0	1.4	± 5.0
QC Mode	RBC	$\times 10^{12}/L$	2.24	± 0.18	4.23	± 0.24	5.17	± 0.30
	HGB	g/dL	6.0	± 0.4	12.2	± 0.6	15.8	± 0.8
	HCT	%	19.8	± 1.5	39.5	± 2.0	51.8	± 2.4
	MCV	fL	88.4	± 5.0	93.4	± 5.0	100.2	± 5.0
	MCH	pg	26.8	± 2.5	28.8	± 2.5	30.6	± 2.5
	MCHC	g/dL	30.3	± 3.0	30.9	± 3.0	30.5	± 3.0
	RDW-CV	%	17.4	± 3.0	15.0	± 3.0	14.3	± 3.0
	RDW-SD	fL	56.4	± 6.0	51.5	± 6.0	52.2	± 8.0
	PLT	$\times 10^9/L$	59	± 20	199	± 40	399	± 60
	MPV	fL	11.3	± 3.0	12.0	± 3.0	11.4	± 3.0
	PDW	/	16.4	± 3.0	16.8	± 3.0	16.5	± 3.0
	PCT	%	0.067	± 0.050	0.239	± 0.100	0.455	± 0.200
	P-LCR	%	37.2	± 20.0	39.8	± 10.0	36.3	± 10.0
	P-LCC	$\times 10^9/L$	22	± 12	79	± 24	145	± 50
	IMG#	$\times 10^9/L$	0.05	± 0.39	0.09	± 0.75	0.28	± 2.16
	IMG%	%	1.3	± 10.0	1.2	± 10.0	1.3	± 10.0
	NRBC#	$\times 10^9/L$	0.319	± 0.319	0.344	± 0.344	0.345	± 0.345
	NRBC%	/100WBC	8.59	± 8.59	4.35	± 4.35	1.63	± 1.63

BC-6D

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT MB0326A



2026-05-10

Instrument	Parameter	Low		Normal		High	
		LOT	MB0326AL	LOT	MB0326AN	LOT	MB0326AH
BC-6800	WBC	$\times 10^9/L$	3.71 \pm 0.50		7.72 \pm 1.00		20.56 \pm 2.50
BC-6600	Neu#	$\times 10^9/L$	2.08 \pm 0.38		3.88 \pm 0.78		11.04 \pm 2.06
QC Mode	Lym#	$\times 10^9/L$	0.93 \pm 0.30		2.69 \pm 0.62		7.01 \pm 1.65
	Mon#	$\times 10^9/L$	0.13 \pm 0.13		0.38 \pm 0.38		1.01 \pm 1.01
	Eos#	$\times 10^9/L$	0.53 \pm 0.34		0.69 \pm 0.69		1.27 \pm 1.27
	Bas#	$\times 10^9/L$	0.04 \pm 0.19		0.08 \pm 0.37		0.23 \pm 1.05
	Neu%	%	56.0 \pm 10.0		50.2 \pm 10.0		53.7 \pm 10.0
	Lym%	%	25.2 \pm 8.0		34.8 \pm 8.0		34.1 \pm 8.0
	Mon%	%	3.5 \pm 3.5		4.9 \pm 4.9		4.9 \pm 4.9
	Eos%	%	14.2 \pm 9.0		9.0 \pm 9.0		6.2 \pm 6.2
	Bas%	%	1.1 \pm 5.0		1.1 \pm 5.0		1.1 \pm 5.0
	RBC	$\times 10^{12}/L$	2.28 \pm 0.18		4.21 \pm 0.24		5.11 \pm 0.30
	HGB	g/dL	6.1 \pm 0.4		12.2 \pm 0.6		15.6 \pm 0.8
	HCT	%	20.1 \pm 1.5		39.2 \pm 2.0		50.7 \pm 2.4
	MCV	fL	88.2 \pm 5.0		93.2 \pm 5.0		99.3 \pm 5.0
	MCH	pg	26.8 \pm 2.5		29.0 \pm 2.5		30.5 \pm 2.5
	MCHC	g/dL	30.3 \pm 3.0		31.1 \pm 3.0		30.7 \pm 3.0
	RDW-CV	%	16.5 \pm 3.0		14.7 \pm 3.0		14.8 \pm 3.0
	RDW-SD	fL	51.6 \pm 6.0		49.4 \pm 6.0		52.0 \pm 8.0
	PLT	$\times 10^9/L$	56 \pm 20		199 \pm 40		412 \pm 60
	MPV	fL	10.9 \pm 3.0		11.7 \pm 3.0		11.1 \pm 3.0
	PDW	/	16.0 \pm 3.0		16.6 \pm 3.0		16.4 \pm 3.0
	PCT	%	0.061 \pm 0.050		0.233 \pm 0.100		0.457 \pm 0.200
	P-LCR	%	34.8 \pm 20.0		37.4 \pm 10.0		33.8 \pm 10.0
	P-LCC	$\times 10^9/L$	19 \pm 12		74 \pm 24		139 \pm 50
	IMG#*	$\times 10^9/L$	0.05 \pm 0.39		0.10 \pm 0.77		0.27 \pm 2.08
	IMG%*	%	1.3 \pm 10.0		1.3 \pm 10.0		1.3 \pm 10.0
	NRBC#	$\times 10^9/L$	0.364 \pm 0.364		0.384 \pm 0.384		0.432 \pm 0.432
	NRBC%	/100WBC	9.81 \pm 9.81		4.97 \pm 4.97		2.10 \pm 2.10

*Software version 1.09 or higher.

BC-6D

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT MB0326A

 2026-05-10

Instrument	Parameter	Low		Normal		High	
		LOT	MB0326AL	LOT	MB0326AN	LOT	MB0326AH
BC-6800Plus	WBC	$\times 10^9/L$	3.72 \pm 0.50		7.79 \pm 1.00		20.99 \pm 2.50
BC-6700Plus	Neu#	$\times 10^9/L$	2.07 \pm 0.38		3.90 \pm 0.78		11.19 \pm 2.10
BC-6600Plus	Lym#	$\times 10^9/L$	0.94 \pm 0.30		2.71 \pm 0.63		7.24 \pm 1.68
QC Mode	Mon#	$\times 10^9/L$	0.13 \pm 0.13		0.38 \pm 0.38		0.97 \pm 0.97
	Eos#	$\times 10^9/L$	0.52 \pm 0.34		0.68 \pm 0.55		1.28 \pm 1.28
	Bas#	$\times 10^9/L$	0.06 \pm 0.06		0.12 \pm 0.12		0.31 \pm 0.31
	Neu%	%	55.5 \pm 10.0		50.1 \pm 10.0		53.3 \pm 10.0
	Lym%	%	25.4 \pm 8.0		34.8 \pm 8.0		34.5 \pm 8.0
	Mon%	%	3.6 \pm 3.6		4.9 \pm 4.9		4.6 \pm 4.6
	Eos%	%	14.0 \pm 9.0		8.7 \pm 7.0		6.1 \pm 6.1
	Bas%	%	1.5 \pm 1.5		1.5 \pm 1.5		1.5 \pm 1.5
	RBC	$\times 10^{12}/L$	2.28 \pm 0.18		4.22 \pm 0.24		5.12 \pm 0.30
	HGB	g/dL	6.2 \pm 0.4		12.2 \pm 0.6		15.7 \pm 0.8
	HCT	%	20.0 \pm 1.5		39.3 \pm 2.0		50.8 \pm 2.4
	MCV	fL	87.9 \pm 5.0		93.1 \pm 5.0		99.3 \pm 5.0
	MCH	pg	27.2 \pm 2.5		28.9 \pm 2.5		30.7 \pm 2.5
	MCHC	g/dL	30.9 \pm 3.0		31.1 \pm 3.0		30.9 \pm 3.0
	RDW-CV	%	16.4 \pm 3.0		14.9 \pm 3.0		14.5 \pm 3.0
	RDW-SD	fL	51.5 \pm 6.0		49.5 \pm 6.0		51.1 \pm 8.0
	PLT	$\times 10^9/L$	56 \pm 20		198 \pm 40		407 \pm 60
	PLT-H	$\times 10^9/L$	59 \pm 20		199 \pm 40		415 \pm 60
	MPV	fL	11.0 \pm 3.0		11.7 \pm 3.0		11.2 \pm 3.0
	PDW	/	16.0 \pm 3.0		16.8 \pm 3.0		16.4 \pm 3.0
	PCT	%	0.062 \pm 0.050		0.232 \pm 0.100		0.456 \pm 0.200
	P-LCR	%	34.5 \pm 20.0		38.3 \pm 10.0		34.3 \pm 10.0
	P-LCC	$\times 10^9/L$	19 \pm 12		76 \pm 24		140 \pm 50
	IMG#	$\times 10^9/L$	0.05 \pm 0.39		0.10 \pm 0.77		0.25 \pm 2.09
	IMG%	%	1.3 \pm 10.0		1.3 \pm 10.0		1.2 \pm 10.0
	NRBC#	$\times 10^9/L$	0.321 \pm 0.321		0.339 \pm 0.339		0.378 \pm 0.378
	NRBC%	/100WBC	8.63 \pm 8.63		4.35 \pm 4.35		1.80 \pm 1.80
	Micro%	%	2.9 \pm 15.0		2.5 \pm 4.0		2.1 \pm 2.5
	Macro%	%	4.5 \pm 3.0		4.5 \pm 5.0		5.3 \pm 10.0
	Neu-X	/	380.7 \pm 120.0		386.0 \pm 120.0		383.5 \pm 120.0
	Neu-Y	/	512.1 \pm 120.0		506.6 \pm 150.0		506.5 \pm 150.0
	Mon-X	/	206.5 \pm 120.0		210.3 \pm 120.0		211.1 \pm 120.0
	PLT-I	$\times 10^9/L$	56 \pm 20		198 \pm 40		407 \pm 60
	ESR*	mm/h	36.52 \pm 10.00		15.49 \pm 4.00		/ \pm /

*The ESR is applicable to software version 1.25 or higher, but not applicable for High level.

BC-6D

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT MB0326A



2026-05-10

Instrument	Parameter	Low		Normal		High	
		LOT	MB0326AL	LOT	MB0326AN	LOT	MB0326AH
BC-6000	WBC	$\times 10^9/L$	3.66 \pm 0.50		7.90 \pm 1.00		21.03 \pm 2.50
BC-6100	Neu#	$\times 10^9/L$	2.06 \pm 0.37		3.95 \pm 0.79		11.31 \pm 2.11
BC-6200	Lym#	$\times 10^9/L$	0.91 \pm 0.30		2.73 \pm 0.64		7.13 \pm 1.69
Counter 60 AL	Mon#	$\times 10^9/L$	0.13 \pm 0.13		0.39 \pm 0.39		0.99 \pm 0.99
BC-6000Plus	Eos#	$\times 10^9/L$	0.51 \pm 0.30		0.71 \pm 0.56		1.28 \pm 1.28
BC-6100Plus	Bas#	$\times 10^9/L$	0.05 \pm 0.05		0.12 \pm 0.12		0.32 \pm 0.32
QC Mode	Neu%	%	56.1 \pm 10.0		50.1 \pm 10.0		53.8 \pm 10.0
	Lym%	%	24.9 \pm 8.0		34.5 \pm 8.0		33.9 \pm 8.0
	Mon%	%	3.5 \pm 3.5		4.9 \pm 4.9		4.7 \pm 4.7
	Eos%	%	14.0 \pm 8.0		9.0 \pm 7.0		6.1 \pm 6.1
	Bas%	%	1.5 \pm 1.5		1.5 \pm 1.5		1.5 \pm 1.5
	RBC	$\times 10^{12}/L$	2.29 \pm 0.18		4.25 \pm 0.24		5.17 \pm 0.30
	HGB	g/dL	6.2 \pm 0.4		12.2 \pm 0.6		15.8 \pm 0.8
	HCT	%	20.7 \pm 1.5		40.0 \pm 2.0		51.2 \pm 2.4
	MCV	fL	90.2 \pm 5.0		94.1 \pm 5.0		99.1 \pm 5.0
	MCH	pg	27.1 \pm 2.5		28.7 \pm 2.5		30.6 \pm 2.5
	MCHC	g/dL	30.0 \pm 3.0		30.5 \pm 3.0		30.8 \pm 3.0
	RDW-CV	%	17.0 \pm 3.0		15.1 \pm 3.0		14.6 \pm 3.0
	RDW-SD	fL	55.8 \pm 6.0		51.8 \pm 6.0		53.1 \pm 8.0
	PLT	$\times 10^9/L$	58 \pm 20		200 \pm 40		419 \pm 60
	MPV	fL	12.2 \pm 3.0		12.6 \pm 3.0		11.6 \pm 3.0
	PDW	/	16.2 \pm 3.0		16.6 \pm 3.0		16.2 \pm 3.0
	PCT	%	0.071 \pm 0.050		0.252 \pm 0.100		0.486 \pm 0.200
	P-LCR	%	42.1 \pm 20.0		43.1 \pm 10.0		37.3 \pm 10.0
	P-LCC	$\times 10^9/L$	24 \pm 12		86 \pm 24		156 \pm 50
	IMG#	$\times 10^9/L$	0.05 \pm 0.39		0.10 \pm 0.77		0.27 \pm 2.08
	IMG%	%	1.3 \pm 10.0		1.3 \pm 10.0		1.3 \pm 10.0
	NRBC#	$\times 10^9/L$	0.317 \pm 0.317		0.350 \pm 0.350		0.334 \pm 0.334
	NRBC%	/100WBC	8.66 \pm 8.66		4.43 \pm 4.43		1.59 \pm 1.59
	Micro%	%	3.0 \pm 15.0		2.5 \pm 4.0		2.1 \pm 2.5
	Macro%	%	4.9 \pm 3.0		5.5 \pm 5.0		6.6 \pm 10.0
	Neu-X	/	381.7 \pm 120.0		408.1 \pm 120.0		402.9 \pm 120.0
	Neu-Y	/	518.4 \pm 120.0		515.8 \pm 150.0		515.6 \pm 150.0
	Mon-X	/	207.3 \pm 120.0		211.5 \pm 120.0		210.1 \pm 120.0
	PLT-I	$\times 10^9/L$	58 \pm 20		200 \pm 40		419 \pm 60

BC-6D

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT MB0326A



2026-05-10

Instrument	Parameter		Low		Normal		High	
			LOT	MB0326AL	LOT	MB0326AN	LOT	MB0326AH
BC-700[B]	WBC	$\times 10^9/L$	3.71	± 0.50	8.05	± 1.00	21.72	± 2.50
BC-760[B]	Neu#	$\times 10^9/L$	2.17	± 0.38	4.24	± 0.81	12.24	± 2.18
BC-700[B] CS	Lym#	$\times 10^9/L$	0.92	± 0.30	2.80	± 0.65	7.38	± 1.74
BC-760[B] CS	Mon#	$\times 10^9/L$	0.13	± 0.13	0.39	± 0.39	1.00	± 1.00
BC-700[R]	Eos#	$\times 10^9/L$	0.44	± 0.44	0.51	± 0.51	0.80	± 0.80
BC-720[R]	Bas#	$\times 10^9/L$	0.05	± 0.18	0.11	± 0.40	0.30	± 1.08
BC-760[R]	Neu%	%	58.4	± 10.0	52.7	± 10.0	56.3	± 10.0
BC-700[R] CS	Lym%	%	24.9	± 8.0	34.8	± 8.0	34.0	± 8.0
BC-760[R] CS	Mon%	%	3.5	± 3.5	4.8	± 4.8	4.6	± 4.6
BP 200n	Eos%	%	11.8	± 11.8	6.3	± 6.3	3.7	± 3.7
BP 260n	Bas%	%	1.4	± 5.0	1.4	± 5.0	1.4	± 5.0
QC Mode	RBC	$\times 10^{12}/L$	2.27	± 0.18	4.24	± 0.24	5.19	± 0.30
	HGB	g/dL	6.1	± 0.4	12.1	± 0.6	15.7	± 0.8
	HCT	%	20.4	± 1.5	40.2	± 2.0	52.6	± 2.4
	MCV	fL	89.7	± 5.0	94.7	± 5.0	101.4	± 5.0
	MCH	pg	26.9	± 2.5	28.5	± 2.5	30.3	± 2.5
	MCHC	g/dL	30.0	± 3.0	30.1	± 3.0	29.8	± 3.0
	RDW-CV	%	17.3	± 3.0	15.0	± 3.0	14.3	± 3.0
	RDW-SD	fL	57.2	± 6.0	52.4	± 6.0	53.0	± 8.0
	PLT	$\times 10^9/L$	61	± 20	201	± 40	410	± 60
	PLT-H	$\times 10^9/L$	61	± 20	204	± 40	415	± 60
	MPV	fL	11.3	± 3.0	12.0	± 3.0	11.3	± 3.0
	PDW	/	16.4	± 3.0	16.9	± 3.0	16.5	± 3.0
	PCT	%	0.069	± 0.050	0.241	± 0.100	0.463	± 0.200
	P-LCR	%	38.0	± 20.0	39.9	± 10.0	36.2	± 10.0
	P-LCC	$\times 10^9/L$	23	± 12	80	± 24	148	± 50
	IMG#	$\times 10^9/L$	0.04	± 0.34	0.10	± 0.84	0.28	± 2.16
	IMG%	%	1.2	± 10.0	1.2	± 10.0	1.3	± 10.0
	NRBC#	$\times 10^9/L$	0.084	± 0.084	0.163	± 0.163	0.480	± 0.480
	NRBC%	/100WBC	2.27	± 2.27	2.02	± 2.02	2.21	± 2.21
	IPF	%	2.5	± 5.0	2.5	± 5.0	2.5	± 5.0
	ESR	mm/h	37.55	± 10.00	14.52	± 4.00	/	± /

BC-6D

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT MB0326A



2026-05-10


Instrument	Parameter		Low		Normal		High	
			LOT	MB0326AL	LOT	MB0326AN	LOT	MB0326AH
BC-7600	WBC	$\times 10^9/L$	3.73	± 0.80	7.96	± 1.00	21.46	± 2.50
BC-7800	Neu#	$\times 10^9/L$	2.08	± 0.50	4.03	± 0.80	11.57	± 2.00
BC-7900	Lym#	$\times 10^9/L$	0.93	± 0.40	2.72	± 0.80	7.25	± 1.60
(Not applicable to Russia)	Mon#	$\times 10^9/L$	0.13	± 0.12	0.38	± 0.38	0.99	± 0.80
	Eos#	$\times 10^9/L$	0.53	± 0.30	0.72	± 0.50	1.35	± 1.20
	Bas#	$\times 10^9/L$	0.06	± 0.10	0.11	± 0.10	0.30	± 0.29
	Neu%	%	56.0	± 12.0	50.5	± 10.0	53.9	± 10.0
	Lym%	%	24.9	± 9.5	34.2	± 10.0	33.8	± 8.0
	Mon%	%	3.5	± 3.0	4.8	± 4.0	4.6	± 3.5
	Eos%	%	14.1	± 7.0	9.1	± 6.0	6.3	± 6.0
	Bas%	%	1.5	± 1.0	1.4	± 1.0	1.4	± 1.5
	RBC	$\times 10^{12}/L$	2.23	± 0.18	4.20	± 0.24	5.13	± 0.30
	HGB	g/dL	6.1	± 0.4	12.2	± 0.6	15.8	± 0.8
	HCT	%	20.0	± 2.0	39.8	± 2.5	52.0	± 3.0
	MCV	fL	89.6	± 5.0	94.8	± 5.0	101.3	± 5.0
	MCH	pg	27.4	± 2.5	29.0	± 2.5	30.8	± 2.5
	MCHC	g/dL	30.5	± 3.0	30.6	± 3.0	30.4	± 3.0
	RDW-CV	%	17.5	± 5.0	15.0	± 5.0	14.3	± 6.0
	RDW-SD	fL	56.5	± 10.0	51.3	± 10.0	52.0	± 12.0
	PLT	$\times 10^9/L$	57	± 20	197	± 40	397	± 60
	PLT-H	$\times 10^9/L$	57	± 20	196	± 40	395	± 60
	MPV	fL	11.1	± 3.0	11.9	± 3.0	11.2	± 3.0
	PDW	/	16.4	± 5.0	16.9	± 5.0	16.5	± 5.0
	PCT	%	0.063	± 0.050	0.234	± 0.100	0.445	± 0.200
	P-LCR	%	36.2	± 10.0	39.5	± 10.0	35.2	± 10.0
	P-LCC	$\times 10^9/L$	21	± 8	78	± 24	140	± 50
	IMG#	$\times 10^9/L$	0.05	± 0.50	0.10	± 0.90	0.26	± 2.20
	IMG%	%	1.3	± 10.0	1.3	± 10.0	1.2	± 10.0
	NRBC#	$\times 10^9/L$	0.322	± 0.300	0.347	± 0.500	0.346	± 0.300
	NRBC%	/100WBC	8.62	± 10.00	4.36	± 10.00	1.61	± 10.00
	Micro%	%	2.6	± 15.0	2.3	± 4.0	2.0	± 2.5
	Macro%	%	4.6	± 3.0	4.9	± 5.0	5.6	± 10.0
	Neu-X	/	396.5	± 120.0	404.5	± 120.0	392.0	± 120.0
	Neu-Y	/	522.9	± 120.0	521.8	± 150.0	518.8	± 150.0
	Mon-X	/	221.0	± 120.0	221.2	± 120.0	220.3	± 120.0
	PLT-I	$\times 10^9/L$	57	± 20	197	± 40	397	± 60
	ESR	mm/h	37.02	± 10.00	13.63	± 4.00	/	$\pm /$
	WBC-M*	$\times 10^9/L$	3.83	± 0.80	8.01	± 1.00	21.62	± 2.50

The ESR is not applicable for High level, and only ESR is involved in BP 200n and BP 260n.

WBC-M applies to BC-7900.

*Research use only.


AS908 Rev.08




BC-6000series_MB0326AL



BC-6000series_MB0326AN




BC-6000series_MB0326AH



BC-6800PLUS_MB0326AL



BC-6800PLUS_MB0326AN



BC-6800PLUS_MB0326AH



BC-6800series_MB0326AL



BC-6800series_MB0326AN



BC-6800series_MB0326AH



BC-700series_MB0326AL



BC-700series_MB0326AN



BC-700series_MB0326AH




BC-7500series_MB0326AL




BC-7500series_MB0326AN




BC-7500series_MB0326AH



BC-7600series_MB0326AL



BC-7600series_MB0326AN



BC-7600series_MB0326AH



BP200nseries_MB0326AL



BP200nseries_MB0326AN



BP200nseries_MB0326AH