

Bion Calibrator Data Sheet For Mindray

				Bion Calibrator		LOT		J601		2028.03.18				
				Biorad		89771/89772		89771/89772		2027.07.31				
제조사	검사항목	Unit	Method	Bion Cal Conc.	Biorad 1				Biorad 2					
					Mean	SD	Range		Mean	SD	Range			
Bion	ALB	g/dL	BCG	3.6	3.7	0.2	3.3	~	4.1	2.6	0.2	2.2	~	3.0
	ALP	U/L	PNPP/AMP/IFCC	236	128	10	108	~	148	541	25	491	~	591
	AMY	U/L	EPS/IFCC	186	67	4	59	~	75	586	20	546	~	626
	AST(GOT)	U/L	UV without P5P(IFCC)	109	51	4	43	~	59	234	10	214	~	254
	ALT(GPT)	U/L	UV without P5P(IFCC)	117	33	3	27	~	39	116	8	100	~	132
	BUN	mg/dL	Urease GLDH	48.6	16.3	1.5	13.3	~	19.3	46.9	3.5	39.9	~	53.9
	CA	mg/dL	Arsenazo 3	9.8	8.9	0.8	7.3	~	10.5	11.7	1.0	9.7	~	13.7
	CHO	mg/dL	COD/POD	167	242	12	218	~	266	106	6	94	~	118
	CK	U/L	UV Rate	375	150	11	129	~	171	450	20	410	~	490
	CREA	mg/dL	Jaffe	3.93	2.31	0.20	1.91	~	2.71	5.56	0.60	4.36	~	6.76
	D-BIL	mg/dL	Oxidation	2.2	0.44	0.10	0.24	~	0.64	1.74	0.20	1.34	~	2.14
	GGT	U/L	Glupa-C IFCC	95	63	6	51	~	75	179	9	161	~	197
	GLU	mg/dL	Hexokinase	204	74	5	64	~	84	292	15	262	~	322
	GLU 1STEP	mg/dL	GOD/POD	190	73	5	63		83	278	15	248		308
	HDL	mg/dL	Direct	56	80	5	70	~	90	23	3	17	~	29
	IP	mg/dL	Molybdate	5.51	3.15	0.30	2.55	~	3.75	6.92	0.80	5.32	~	8.52
	LDH	U/L	LD(P>L) Tris	558	379	17.5	344	~	414	931	40	851	~	1011
	LDL	mg/dL	Direct	65	106	7	92	~	120	50	4	42	~	58
	T-BIL	mg/dL	Oxidation	4.3	1.14	0.12	0.90	~	1.38	4.01	0.20	3.61	~	4.41
	TG	mg/dL	Enzymetic	145	167	10	147	~	187	80	5	70	~	90
TP	g/dL	Biuret	5.6	6.4	0.35	5.7	~	7.1	4.4	0.2	4.0	~	4.8	
UA	mg/dL	효소법	5.49	4.46	0.35	3.76	~	5.16	9.41	0.7	8.01	~	10.81	

JWBS	CA	mg/dL	Arsenazo 3	9.8	8.9	0.8	7.3	~	10.5	11.7	1.0	9.7	~	13.7
	CK	U/L	UV Rate	375	150	10.5	129	~	171	450	20	410	~	490
	IP	mg/dL	Molybdate	5.51	3.15	0.30	2.55	~	3.75	6.92	0.80	5.32	~	8.52
	MG	mg/dL	Xylidyl Blue	2.9	2.3	0.3	1.7	~	2.9	4.9	0.4	4.1	~	5.7

*본 Data는 기종에 따라 오차가 발생할 수도 있음을 알려드립니다.